

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	30.0 °C	-0.062 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	31.8 %HR	1.230 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	60.018 %HR	61.8 %HR	1.782 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.550 %HR	70.8 %HR	0.250 %HR	± 3.0 %HR	± 1.9E+00 %HR