

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	29.9 °C	-0.162 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	30.4 %HR	-0.170 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	60.018 %HR	60.7 %HR	0.682 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.550 %HR	69.2 %HR	-1.350 %HR	± 3.0 %HR	± 1.9E+00 %HR