

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	29.9 °C	-0.162 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	30.7 %HR	0.130 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	60.018 %HR	59.5 %HR	-0.518 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.550 %HR	69.2 %HR	-1.350 %HR	± 3.0 %HR	± 1.9E+00 %HR