

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
3599.561 s	3599.56 s	-0.001 s	0.000%	± 0.0012 %	± 5.6E-02 s
3599.223 s	3599.22 s	-0.003 s	0.000%	± 0.0012 %	± 5.6E-02 s
3599.421 s	3599.42 s	-0.001 s	0.000%	± 0.0012 %	± 5.6E-02 s