

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
60.561 s	60.56 s	-0.001 s	-0.002%	± 0.0012 %	±	7.0E-02 s
30.223 s	30.22 s	-0.003 s	-0.010%	± 0.0012 %	±	7.0E-02 s
90.421 s	90.42 s	-0.001 s	-0.001%	± 0.0012 %	±	7.0E-02 s