

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
60 RPM	60.01 RPM	0.01 RPM	± 0.038 RPM	±	7.0E-03 RPM
600 RPM	600.1 RPM	0.1 RPM	± 0.38 RPM	±	5.8E-02 RPM
6000 RPM	6000 RPM	0 RPM	± 3.8 RPM	±	5.8E-02 RPM
60000 RPM	60000 RPM	0 RPM	± 20 RPM	±	8.0E-01 RPM
99960 RPM	99958 RPM	-2 RPM	± 32 RPM	±	8.4E-01 RPM