

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	29.9 °C	-0.162 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	32.0 %HR	1.430 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.100 %HR	53.0 %HR	2.900 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.550 %HR	74.0 %HR	3.450 %HR	± 5.0 %HR	± 1.9E+00 %HR