

Tipo de Medida <i>(Measurement Type)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
Exterior / (Exterior)	# 0 mm	0.01 mm	10 µm	± 30 µm	±	7.4E+00 µm
Exterior / (Exterior)	41.30 mm	41.32 mm	20 µm	± 30 µm	±	7.4E+00 µm
Exterior / (Exterior)	131.40 mm	131.42 mm	20 µm	± 30 µm	±	7.4E+00 µm
Exterior / (Exterior)	150.00 mm	150.02 mm	20 µm	± 30 µm	±	7.4E+00 µm
Exterior / (Exterior)	200.00 mm	200.02 mm	20 µm	± 30 µm	±	7.4E+00 µm
Interior / (Interior)	20.00 mm	20.01 mm	10 µm	± 30 µm	±	8.2E+00 µm
Interior / (Interior)	30.00 mm	30.02 mm	20 µm	± 30 µm	±	8.2E+00 µm
Profundidad / (Depth)	# 0 mm	0.01 mm	10 µm	± 30 µm	±	7.1E+00 µm