

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.522 °C	15.4 °C	-0.122 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.1 °C	-0.149 °C	± 0.5 °C	± 4.0E-01 °C
50 °C	50.218 °C	50.1 °C	-0.118 °C	± 0.5 °C	± 1.1E+00 °C
25 %HR	25.458 %HR	27 %HR	1.342 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.008 %HR	50 %HR	0.292 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	76 %HR	0.440 %HR	± 3.0 %HR	± 1.9E+00 %HR