

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
60 RPM	59.9 RPM	-0.1 RPM	± 0.13 RPM	±	5.8E-02 RPM
600 RPM	600.1 RPM	0.1 RPM	± 0.40 RPM	±	5.8E-02 RPM
6000 RPM	6000 RPM	0 RPM	± 4.0 RPM	±	5.8E-01 RPM
60000 RPM	60000 RPM	0 RPM	± 30 RPM	±	8.0E-01 RPM
99960 RPM	99961 RPM	1 RPM	± 50 RPM	±	8.4E-01 RPM