

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.522 °C	15.4 °C	-0.122 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.249 °C	24.7 °C	-0.549 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.218 °C	49.8 °C	-0.418 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.458 %HR	23.0 %HR	-2.458 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.008 %HR	48.0 %HR	-2.008 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	74.0 %HR	-1.560 %HR	± 5.0 %HR	± 1.9E+00 %HR