

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.522 °C	15.1 °C	-0.422 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.249 °C	24.8 °C	-0.449 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.218 °C	49.7 °C	-0.518 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.458 %HR	22 %HR	-3.458 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.008 %HR	42 %HR	-8.008 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	67 %HR	-8.560 %HR	± 5.0 %HR	± 1.9E+00 %HR