

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.01345 ml	10 ml	-0.01345 ml	± 0.25 ml	±	3.3E-01 µl
29.95791 ml	30 ml	0.04209 ml	± 0.25 ml	±	9.6E-01 µl
49.99303 ml	50 ml	0.00697 ml	± 0.25 ml	±	1.6E+00 µl