

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.02645 ml	10 ml	-0.02645 ml	± 0.5 ml	±	3.3E-01 µl
49.98541 ml	50 ml	0.01459 ml	± 0.5 ml	±	1.6E+00 µl
100.06328 ml	100 ml	-0.06328 ml	± 0.5 ml	±	3.2E+00 µl