

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
29.93074 ml	30 ml	0.06926 ml	± 1 ml	±	9.6E-01 µl
149.89687 ml	150 ml	0.10313 ml	± 1 ml	±	4.8E+00 µl
250.15569 ml	250 ml	-0.15569 ml	± 1 ml	±	8.0E+00 µl