

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.00607 ml	50 ml	-0.00607 ml	± 2 ml	±	1.6E+00 µl
249.95984 ml	250 ml	0.04016 ml	± 2 ml	±	8.0E+00 µl
500.14792 ml	500 ml	-0.14792 ml	± 2 ml	±	1.6E+01 µl