

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.02477 ml	100 ml	-0.02477 ml	± 3 ml	±	3.2E+00 µl
500.06409 ml	500 ml	-0.06409 ml	± 3 ml	±	1.6E+01 µl
1,000.01666 ml	1000 ml	-0.01666 ml	± 3 ml	±	3.2E+01 µl