

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.0037 µl	100 µl	-0.0037 µl	± 0.5 µl	±	5.3E-02 µl
500.2190 µl	500 µl	-0.2190 µl	± 0.5 µl	±	5.9E-02 µl
1,000.2174 µl	1000 µl	-0.2174 µl	± 0.5 µl	±	7.8E-02 µl