

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	31.1 °C	1.038 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	32.0 %HR	1.430 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.008 %HR	54.0 %HR	3.992 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	70.448 %HR	74.0 %HR	3.552 %HR	± 5.0 %HR	± 1.9E+00 %HR