

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
in	in	in	in	(Uncertainty)	in	in	in	in	(Uncertainty)
0 a 1	1.00082	-0.00082	± 0.03	± 1.4E-03 in					
0 a 1.5	1.50048	-0.00048	± 0.5	± 1.4E-03 in					
0 a 2	2.00062	-0.00062	± 0.6	± 1.4E-03 in					
0 a 2.5	2.50063	-0.00063	± 0.7	± 1.4E-03 in					
0 a 3	3.00081	-0.00081	± 0.8	± 1.4E-03 in					
0 a 4	4.00077	-0.00077	± 1	± 1.4E-03 in					
0 a 4.5	4.50092	-0.00092	± 1.1	± 1.4E-03 in					
0 a 5	5.00083	-0.00083	± 1.2	± 1.4E-03 in					
0 a 5.5	5.50076	-0.00076	± 1.3	± 1.4E-03 in					
0 a 6	6.00089	-0.00089	± 1.4	± 1.4E-03 in					