

Patrón (Standard)		UBP (UIT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
30 Hz	1.000 g-pk	1.001 g-pk	0.00 g-pk	± 0.01 g-pk	±	0.021 g-pk
50 Hz	1.000 g-pk	0.999 g-pk	0.00 g-pk	± 0.01 g-pk	±	0.020 g-pk
80 Hz	1.000 g-pk	0.998 g-pk	0.00 g-pk	± 0.01 g-pk	±	0.020 g-pk
100 Hz	1.000 g-pk	1.002 g-pk	0.00 g-pk	± 0.01 g-pk	±	0.021 g-pk
500 Hz	1.000 g-pk	1 g-pk	0.00 g-pk	± 0.01 g-pk	±	0.020 g-pk
1000 Hz	1.000 g-pk	1.00 g-pk	0.00 g-pk	± 0.01 g-pk	±	0.021 g-pk