

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.042 °C	29.6 °C	-0.442 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.333 %HR	33.2 %HR	3.867 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.824 %HR	53.1 %HR	3.276 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	73.392 %HR	72.5 %HR	-0.892 %HR	± 3.0 %HR	± 1.9E+00 %HR