

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.042 °C	29.5 °C	-0.542 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.333 %HR	33.0 %HR	3.667 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.824 %HR	52.8 %HR	2.976 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	73.392 %HR	74.6 %HR	1.208 %HR	± 3.0 %HR	± 1.9E+00 %HR