

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.310 °C        | -25.3 °C  | 0.010 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.121 °C         | -0.4 °C   | -0.279 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 8 °C              | 8.102 °C          | 8.0 °C    | -0.102 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |