

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.310 °C	-25.3 °C	0.010 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.121 °C	0.0 °C	0.121 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.102 °C	8.2 °C	0.098 °C	± 0.5 °C	± 4.0E-01 °C