

| Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |              |
|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|--------------|
| 50.0 mmHg                  | 50 mmHg            | 0.0 mmHg               | ± 5 mmHg                      | ±                                    | 6.6E-01 mmHg |
| 101.0 mmHg                 | 100 mmHg           | -1.0 mmHg              | ± 5 mmHg                      | ±                                    | 6.5E-01 mmHg |
| 150.0 mmHg                 | 150 mmHg           | 0.0 mmHg               | ± 5 mmHg                      | ±                                    | 6.6E-01 mmHg |
| 201.0 mmHg                 | 200 mmHg           | -1.0 mmHg              | ± 5 mmHg                      | ±                                    | 6.8E-01 mmHg |
| 251.0 mmHg                 | 250 mmHg           | -1.0 mmHg              | ± 5 mmHg                      | ±                                    | 7.5E-01 mmHg |
| 299.0 mmHg                 | 300 mmHg           | 1.0 mmHg               | ± 5 mmHg                      | ±                                    | 7.5E-01 mmHg |