

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)	
51.0 mmHg	50 mmHg	-1.0 mmHg	± 5 mmHg	±	6.6E-01 mmHg
101.0 mmHg	100 mmHg	-1.0 mmHg	± 5 mmHg	±	6.5E-01 mmHg
149.0 mmHg	150 mmHg	1.0 mmHg	± 5 mmHg	±	6.6E-01 mmHg
199.0 mmHg	200 mmHg	1.0 mmHg	± 5 mmHg	±	6.8E-01 mmHg
249.0 mmHg	250 mmHg	1.0 mmHg	± 5 mmHg	±	7.5E-01 mmHg
299.0 mmHg	300 mmHg	1.0 mmHg	± 5 mmHg	±	7.5E-01 mmHg