

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.8	0.799000	1.270	± 1.8	± 1.5E+00 mm					
0 a 0.9	0.898000	2.280	± 1.9	± 1.5E+00 mm					
0 a 1	1.002000	-1.720	± 2	± 1.5E+00 mm					
0 a 1.1	1.099000	1.280	± 2	± 1.5E+00 mm					
0 a 1.2	1.198000	2.290	± 2.1	± 1.5E+00 mm					
0 a 1.3	1.299000	1.290	± 2.2	± 1.5E+00 mm					
0 a 1.4	1.398000	2.300	± 2.3	± 1.5E+00 mm					
0 a 1.5	1.498000	2.300	± 2.4	± 1.5E+00 mm					
0 a 1.6	1.598000	2.310	± 2.4	± 1.5E+00 mm					
0 a 1.7	1.698000	2.310	± 2.5	± 1.5E+00 mm					
0 a 1.8	1.801000	-0.680	± 2.6	± 1.5E+00 mm					
0 a 1.9	1.898000	2.330	± 2.7	± 1.5E+00 mm					
0 a 2.1	2.098000	2.340	± 2.8	± 1.5E+00 mm					