

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1005 RPM	1000 RPM	-5 RPM	± 20 RPM	±	8.2E-02 RPM
1254 RPM	1250 RPM	-4 RPM	± 20 RPM	±	1.1E-01 RPM
1502 RPM	1500 RPM	-2 RPM	± 20 RPM	±	1.1E-01 RPM
1758 RPM	1750 RPM	-8 RPM	± 20 RPM	±	1.1E-01 RPM
2006 RPM	2000 RPM	-6 RPM	± 20 RPM	±	1.1E-01 RPM