

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.050 °C	0.3 °C	0.250 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.4 °C	0.407 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.3 °C	0.293 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	31.3 %HR	0.934 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.396 %HR	50.6 %HR	0.204 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.039 %HR	70.3 %HR	0.261 %HR	± 3.0 %HR	± 1.9E+00 %HR