

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.083 °C	12.3 °C	0.217 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.034 °C	27.2 °C	0.166 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.095 °C	42.1 °C	0.005 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.441 %HR	54.9 %HR	2.459 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	71.981 %HR	72.2 %HR	0.219 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.125 %HR	91.3 %HR	-0.825 %HR	± 3.0 %HR	± 1.9E+00 %HR