

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
12 °C	12.083 °C	12.3 °C	0.217 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.034 °C	26.9 °C	-0.134 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.095 °C	41.9 °C	-0.195 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.441 %HR	53.5 %HR	1.059 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	71.981 %HR	70.2 %HR	-1.781 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.125 %HR	93.1 %HR	0.975 %HR	± 3.0 %HR	± 1.9E+00 %HR