

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.083 °C	11.8 °C	-0.283 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.034 °C	27.4 °C	0.366 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.095 °C	42.3 °C	0.205 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.441 %HR	55.2 %HR	2.759 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	71.981 %HR	73.0 %HR	1.019 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.125 %HR	91.4 %HR	-0.725 %HR	± 3.0 %HR	± 1.9E+00 %HR