

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0994 ml	0.1 ml	0.0006 ml	± 0.006 ml	±	5.1E-02 µl
0.9992 ml	1 ml	0.0008 ml	± 0.006 ml	±	7.7E-02 µl
10.0002 ml	10 ml	-0.0002 ml	± 0.03 ml	±	5.7E-01 µl
24.995364 ml	25 ml	0.0046 ml	± 0.03 ml	±	1.4E+00 µl