

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	
4.9991 ml	5 ml	0.0009 ml	± 0.01 ml	±	2.9E-01 µl
10.0002 ml	10 ml	-0.0002 ml	± 0.03 ml	±	5.7E-01 µl
20.0023 ml	20 ml	-0.0023 ml	± 0.03 ml	±	1.1E+00 µl
24.995364 ml	25 ml	0.0046 ml	± 0.03 ml	±	1.4E+00 µl