

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
49.98501 ml	50 ml	0.01499 ml	± 0.5 ml	±	1.6E+00 µl
250.06403 ml	250 ml	-0.06403 ml	± 1.0 ml	±	8.0E+00 µl
500.02879 ml	500 ml	-0.02879 ml	± 2.5 ml	±	1.6E+01 µl