

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1,000.0369 µl	1000 µl	-0.0369 µl	± 8 µl	±	7.8E-02 µl
1999.9736 µl	2000 µl	0.0264 µl	± 16 µl	±	1.4E-01 µl
3,000.0106 µl	3000 µl	-0.0106 µl	± 24 µl	±	1.9E-01 µl
3999.9473 µl	4000 µl	0.0527 µl	± 32 µl	±	2.5E-01 µl
5000.1847 µl	5000 µl	-0.1847 µl	± 40 µl	±	3.0E-01 µl