

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.623 °C	16.5 °C	0.877 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.750 °C	26.3 °C	0.550 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.418 °C	50.2 °C	-0.218 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.253 %HR	24.0 %HR	-1.253 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.396 %HR	48.0 %HR	-2.396 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	75.0 %HR	-0.560 %HR	± 5.0 %HR	± 1.9E+00 %HR