

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
19,897.7 lx	19,893 lx	4.7 lx	± 597 lx	±	9.8E+01 lx
1,035.5 lx	1,028 lx	7.5 lx	± 31 lx	±	1.1E+00 lx
340.7 lx	332 lx	8.7 lx	± 10 lx	±	1.8E-01 lx
30.2 lx	29 lx	1.2 lx	± 2 lx	±	1.8E-01 lx