

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	20.005 °C	20.2 °C	0.195 °C	± 0.5 °C	± 1.70E-01 °C
25 °C	25.450 °C	25.6 °C	0.150 °C	± 0.5 °C	± 1.70E-01 °C
30 °C	30.032 °C	30.4 °C	0.368 °C	± 0.5 °C	± 1.60E-01 °C