

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	30.5 °C	0.438 °C	± 0.5 °C	± 1.60E-01 °C
25 %HR	25.560 %HR	24.1 %HR	-1.460 %HR	± 5.0 %HR	± 1.40E+00 %HR
45 %HR	45.499 %HR	42.1 %HR	-3.399 %HR	± 5.0 %HR	± 1.20E+00 %HR
75 %HR	75.049 %HR	70.4 %HR	-4.649 %HR	± 5.0 %HR	± 1.10E+00 %HR