

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.005 °C	20.2 °C	0.195 °C	± 0.5 °C	± 1.70E-01 °C
25 °C	25.450 °C	25.6 °C	0.150 °C	± 0.5 °C	± 1.70E-01 °C
30 °C	30.032 °C	30.4 °C	0.368 °C	± 0.5 °C	± 1.60E-01 °C
25 %HR	25.560 %HR	24.1 %HR	-1.460 %HR	± 5.0 %HR	± 1.40E+00 %HR
45 %HR	45.499 %HR	42.10 %HR	-3.399 %HR	± 5.0 %HR	± 1.00E+00 %HR
75 %HR	75.049 %HR	70.4 %HR	-4.649 %HR	± 5.0 %HR	± 1.20E+00 %HR