

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 20 °C             | 20.005 °C         | 17.5 °C   | -2.505 °C     | ± 0.5 °C             | ± 1.70E-01 °C               |
| 25 °C             | 25.450 °C         | 20.1 °C   | -5.350 °C     | ± 0.5 °C             | ± 1.70E-01 °C               |
| 30 °C             | 30.032 °C         | 25.9 °C   | -4.132 °C     | ± 0.5 °C             | ± 1.60E-01 °C               |