

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.005 °C	20.4 °C	0.395 °C	± 0.5 °C	± 1.70E-01 °C
25 °C	25.450 °C	25.7 °C	0.250 °C	± 0.5 °C	± 1.70E-01 °C
30 °C	30.032 °C	30.4 °C	0.368 °C	± 0.5 °C	± 1.60E-01 °C