

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 20 °C                      | 20.005 °C                  | 20.4 °C            | 0.395 °C               | ± 0.5 °C                      | ± 1.70E-01 °C                        |
| 25 °C                      | 25.450 °C                  | 25.8 °C            | 0.350 °C               | ± 0.5 °C                      | ± 1.70E-01 °C                        |
| 30 °C                      | 30.032 °C                  | 30.4 °C            | 0.368 °C               | ± 0.5 °C                      | ± 1.60E-01 °C                        |