

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	20.005 °C	20.5 °C	0.495 °C	± 0.5 °C	± 1.60E-01 °C
25 °C	25.450 °C	25.9 °C	0.450 °C	± 0.5 °C	± 1.40E+00 °C
30 °C	30.032 °C	30.5 °C	0.468 °C	± 0.5 °C	± 1.20E+00 °C