

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	30.4 °C	0.338 °C	± 0.5 °C	± 1.60E-01 °C
25 %HR	25.560 %HR	24.5 %HR	-1.060 %HR	± 5.0 %HR	± 1.40E+00 %HR
45 %HR	45.499 %HR	43.3 %HR	-2.199 %HR	± 5.0 %HR	± 1.20E+00 %HR
75 %HR	75.049 %HR	72.5 %HR	-2.549 %HR	± 5.0 %HR	± 1.10E+00 %HR