

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.012 °C	29.9 °C	-0.112 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	25.2 %HR	-4.859 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	45.3 %HR	-4.728 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	65.8 %HR	-4.545 %HR	± 3.0 %HR	± 1.9E+00 %HR