

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.012 °C	30.8 °C	0.788 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	24.0 %HR	-6.059 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	41.1 %HR	-8.928 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	58.2 %HR	-12.145 %HR	± 5.0 %HR	± 1.9E+00 %HR