

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
5 °C	5.495 °C	5.1 °C	-0.395 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.028 °C	25.0 °C	-0.028 °C	± 1.0 °C	± 4.0E-01 °C
70 %HR	70.345 %HR	66.0 %HR	-4.345 %HR	± 5.0 %HR	± 1.9E+00 %HR